

EU Declaration of Conformity

Unit type Minicentrifuges-vortexes

Models **FV-2400, FVL-2400N, MSC-3000, MSC-6000, CVP-2**

Serial number 14 digits styled XXXXXYYMMZZZZ, where XXXXXX is model code, YY and MM – year and month of production, ZZZZ – unit number.

Manufacturer SIA BIOSAN
Latvia, LV-1067, Riga, Ratsupites str. 7/2

Applicable Directives EMC Directive 2014/30/EU
LVD Directive 2014/35/EU
RoHS2 2011/65/EU
WEEE 2012/19/EU

Applicable Standards LVS EN 61326-1: 2013
Electrical equipment for measurement, control and laboratory use. EMC requirements. General requirements.
LVS EN 61010-1: 2011
Safety requirements for electrical equipment for measurement, control, and laboratory use. General requirements.
LVS EN 61010-2-020: 2006
Particular requirements for laboratory centrifuges.

We declare that this product conforms to the requirements of the above Directives



Signature

Svetlana Bankovska
Managing director

19.07.2016.

Date



Signature

Aleksandr Shevchik
Engineer of R&D

19.07.2016

Date